

ASCEND

Centrifuge Gel Tubes

Immediately
after draw



After clotting

Minimum 1/2 hour
to maximum 2 hours
after draw



**ACTION
ITEMS**

15 MINUTES, swing bucket centrifuge
20 MINUTES, fixed angle centrifuge