


Centrifuge Gel Tubes

Immediately
after draw



After clotting
Minimum 1/2 hour
to maximum 2 hours
after draw



15 minutes, swing bucket centrifuge
20 minutes, fixed angle centrifuge