

ASCEND Gel Tubes

Centrifuge
immediately
after draw



Centrifuge
after clotting

Minimum 1/2 hour



15 MINUTES, swing bucket centrifuge

20 MINUTES, fixed angle centrifuge

Process all specimens within 2 hours of collection