


ASCEND
Gel Tubes

Centrifuge
immediately
after draw



Centrifuge
after clotting
Minimum 1/2 hour



15 MINUTES, swing bucket centrifuge
20 MINUTES, fixed angle centrifuge

Process all specimens within 2 hours of collection